

LITHIUM ION BATTERY TEST SUMMARY

UN 38.3 Lithium Battery Test Summary for ergoflix® L Battery

1 BATTERY MANUFACTURER

Jiangsu Fengchi Green Power Co., Ltd.
No. 82 Xinzhong Road, Xinzhuang Street,
Yixing City, Jiangsu Province, 214200,
China

2 PRODUCT MANUFACTURER

Nanjing Jin Bai He Medical Apparatur Co., Ltd.
NO.12 Wuge Road, Jiangning District,
Nanjing, Jiangsu, 211113 China
+86 85 78 08 66
info@jbhmedical.com

3 UN38.3 TEST LAB

Shanghai Research Institute Of Chemical Industry Testing Center
Shanghai TECH, Chemical Industry Testing Co., Ltd.
Service Hotline: 02152569800
Tel.: +86 21 52 81 53 77-0713/-0714/-0717/-0718
Fax: +86 21 52 81 53 77-0716/-0745
West entrance, No. 345 East Yun Ling Road, Shanghai, 200062

4 TEST REPORT NUMBER

No. 1116050230

5 DATE OF TEST REPORT

2016-05-20

6 DESCRIPTION OF CELL OR BATTERY

Long, aluminium housing, Model No. ITP 2410
Polymer lithium-ion batteries
24V/6000 mAh / 144 Wh, 1400 g – 1550 g
L: 410mm, W: 65mm, T: 26mm

7 UN38.3 TESTS PERFORMED AND SUCCESSFULLY PASSED

T.1, T.2, T.3, T.4, T.5, T.7. and T.8. (Note that T.6 and T.8 are not applicable to batteries.)

8 ASSEMBLED BATTERY TESTING REQUIREMENTS

Not Applicable

9 EDITION OF UN MANUAL OF TESTS AND CRITERIA USED

ST/SG/AC.10/11/Rev.5/Amend. 1& Amend. 2 38.3 UNITED NATIONS
„Recommendations on the Transport of Dangerous GOODS“
Manual of Tests and Criteria ST/SG/AC.10/11/Rev.5/Amend. 1& Amend. 2 38.3

10 NAME OF SIGNATORY

